

425-487-3157 • email - info@aseptico.com

INSTRUCTIONS FOR USE ARU-01A - Portable X-Ray Stand Extension Arm

- 1. Fully extend front two legs of tripod stand.
- 2. Extend rear leg approximately half way out.
- 3. Attach Extension Arm onto the adjustable center post.
- 4. With assistant, mount X-Ray as shown, use washer and wing nut.
- 5. Take care in making sure all connections are tightened, both on tripod stand and extension arm.
- 6. Carefully check balance of assembled system, ensuring the unit will not easily tip over. Adjust rear leg on tripod stand as necessary.
- 7. Height adjustment to X-Ray head can be achieved by raising and lowering height adjustment on tripod stand.



Half extended

Fully extended :